

**650—29.8(153) Facility and equipment requirements for moderate sedation, deep sedation or general anesthesia.**

**29.8(1)** Change of address or addition of facility location(s). A permit holder shall notify the board office in writing within 60 days of a change in location or the addition of a sedation facility.

**29.8(2)** Facilities shall be permanently equipped. A dentist who administers moderate sedation, deep sedation or general anesthesia in a facility is required to be trained in and maintain, at a minimum, the following equipment to be properly equipped:

- a.* Electrocardiogram (EKG) monitor;
- b.* Positive pressure oxygen;
- c.* Suction;
- d.* Laryngoscope and blades;
- e.* Endotracheal tubes;
- f.* Magill forceps;
- g.* Oral airways;
- h.* Stethoscope;
- i.* Blood pressure monitoring device;
- j.* Pulse oximeter;
- k.* Emergency drugs;
- l.* Defibrillator;
- m.* Capnography machine to monitor end-tidal carbon dioxide;
- n.* Pretracheal or precordial stethoscope; and
- o.* Any additional equipment necessary to establish intravascular or intraosseous access, which shall be available until the patient meets discharge criteria.

**29.8(3)** The board or designated agents of the board may conduct facility inspections. The actual costs associated with the on-site evaluation of the facility shall be the primary responsibility of the licensee. The cost to the licensee shall not exceed the fee specified in 650—Chapter 15.

[ARC 4556C, IAB 7/17/19, effective 8/21/19]